



Republic of the Philippines
AGUSAN DEL SUR STATE COLLEGE OF AGRICULTURE AND TECHNOLOGY
 Bunawan, Agusan del Sur

Office of the Bids and Awards Committee

Supplemental Bid Bulletin No. 01

Project: Procurement of Soil Microbiology Service Laboratory Equipment

This Supplemental Bid Bulletin is issued to all prospective bidders to clarify, modify and/or amend items in the Bidding Documents as discussed and agreed during the Pre-bid Conference held on January 3, 2025 at 10:00 o'clock in the morning at the Board Room, ASSCAT, San Teodoro, Bunawan, Agusan del Sur.

A. Technical Specifications: (Amendments are indicated in Red)

Particulars	Amendments
<p>BIOSAFETY CABINET</p> <ul style="list-style-type: none"> • Airflow system: 100% air exhaust • Number of colonies: ≤0.5pcs/dish hour (Φ90mm culture plate) • Wind speed: Average suction wind speed: ≥0.55±0.025m/s Average descending wind speed: ≥0.3±0.025m/s • Filtration Level / Filtration Efficiency: HEPA of borosilicate glass fiber material: ≥99.995%, @0.3µm Optional ULPA: ≥99.9995% • Noise: ≤65dB(A) • Illumination: ≥800Lux • Vibration half peak: ≤5µm • Power supply: AC single phase 220V/50Hz • Maximum power consumption: 500w • Weight: 210kg • Internal Size (mm) W*D*H: 1040*650*620 • External Size (mm) W*D*H: 1200*800*2270 • High efficiency filter specification and quantity: 980*490*50*1 520*380*70*1 • Specification and quantity of fluorescent lamp/UV lamp: 20W*2/20W*1 	<p>BIOSAFETY CABINET</p> <ul style="list-style-type: none"> • Airflow system: 100% air exhaust • Number of colonies: ≤0.5pcs/dish hour (Φ90mm culture plate) • Wind speed: Average suction wind speed: ≥0.55±0.025m/s Average descending wind speed: ≥0.3±0.025m/s • Filtration Level / Filtration Efficiency: HEPA of borosilicate glass fiber material: ≥99.995%, @0.3µm Optional ULPA: ≥99.9995% • Noise: ≤65dB(A) • Illumination: ≥800Lux • Vibration half peak: ≤5µm • Power supply: AC single phase 220V/50Hz • Maximum power consumption: 500w • Weight: 200-210kg • Internal Size (mm) W*D*H: 1020-1040*620-650*600-620 • External Size (mm) W*D*H: 1200-1210*800-820*2270-2280 • High efficiency filter specification and quantity: 980*490*50*1 520*380*70*1 • Specification and quantity of fluorescent lamp/UV lamp: 20W*2/20W*1



Administration Building, P-1, San Teodoro, Bunawan, Agusan del Sur
 mobile no: +639483679266 | e-mail address: op@asscat.edu.ph
 website: <https://asscat.edu.ph> | Facebook page: <https://www.facebook.com/ASSCATARISE>

Accountability | Responsiveness | Innovativeness | Selflessness | Excellence





Republic of the Philippines
AGUSAN DEL SUR STATE COLLEGE OF AGRICULTURE AND TECHNOLOGY
Bunawan, Agusan del Sur

Office of the Bids and Awards Committee

DIGITAL VERTICAL AUTOCLAVE (75 LITERS) <ul style="list-style-type: none">• LCD display: with LCD display• Chamber volume: 100L(φ440×650) mm• Timer: 0~99min or 0~99hour59min• Power: 4.5KW AC220V.50Hz• G.W/N.W: 105-110Kg/85-90Kg	DIGITAL VERTICAL AUTOCLAVE <ul style="list-style-type: none">• LCD display: with LCD display• Chamber volume: 75-100L• Timer: 0~99min or 0~99hour59min• Power: 4.5KW AC220V.50Hz• G.W/N.W: 105-110Kg/85-90Kg
---	---

For guidance and information of all concerned.


CARMELO S. LLANTO, Ph.D.
BAC Chairperson